

CERTIFICATE OF CONFORMITY




1. **HAZARDOUS LOCATION ELECTRICAL EQUIPMENT PER CANADIAN REQUIREMENTS**
2. **Certificate No:** FM17CA0089X
3. **Equipment:** HV-CEP Series, HW-CEP Series, Platform Scale
(Type Reference and Name)
4. **Name of Listing Company:** A&D Company, Limited
5. **Address of Listing Company:** 1-243, Asahi, Kitamoto-shi, Saitama
Japan
6. The examination and test results are recorded in confidential report number:

3060570 dated 12th January 2018
7. FM Approvals LLC, certifies that the equipment described has been found to comply with the following Approval standards and other documents:

CAN/CSA-C22.2 No. 60079-0:2015, CAN/CSA-C22.2 No. 60079-11:2014,
CAN/CSA-C22.2 No. 61010-1:2012
8. If the sign 'X' is placed after the certificate number, it indicates that the equipment is subject to specific conditions of use specified in the schedule to this certificate.
9. This certificate relates to the design, examination and testing of the products specified herein. The FM Approvals surveillance audit program has further determined that the manufacturing processes and quality control procedures in place are satisfactory to manufacture the product as examined, tested and Approved.
10. **Equipment Ratings:**

Intrinsically Safe for Class I, Division 1, Groups C and D; Ex ia for Class I, Zone 0, Group IIB hazardous (classified) locations, with an ambient temperature rating of T4 for $-25^{\circ}\text{C} \leq T_a \leq +40^{\circ}\text{C}$.
11. The marking of the equipment shall include:

Certificate issued by:



J. E. Marquedant
Manager, Electrical Systems

12 January 2018

Date

To verify the availability of the Approved product, please refer to www.approvalguide.com

THIS CERTIFICATE MAY ONLY BE REPRODUCED IN ITS ENTIRETY AND WITHOUT CHANGE

FM Approvals LLC. 1151 Boston-Providence Turnpike, Norwood, MA 02062 USA
T: +1 (1) 781 762 4300 F: +1 (1) 781 762 9375 E-mail: information@fmapprovals.com www.fmapprovals.com

SCHEDULE



Canadian Certificate Of Conformity No: FM17CA0089X

Class I, Division 1, Groups C and D, T4 for $-25^{\circ}\text{C} \leq T_a \leq +40^{\circ}\text{C}$;
Ex ia for Class I, Zone 0, Group IIB, T4 for $-25^{\circ}\text{C} \leq T_a \leq +40^{\circ}\text{C}$.

12. Description of Equipment:

General – The HV/HW-CEP series are battery-powered platform scales. Each model contains a single PCB, load cell, display, four internal 1.5V alkaline batteries, and a current-limiting protective resistor.

Model variants have different specifications for parameters such as load cell spring material, resolution and range, which are reflected in the maximum weighing capacity.

Construction – The enclosures are constructed of stainless steel and plastic, with the base being over 50% aluminum. The enclosures are of three different sizes (Small, Medium, and Large), based on the weighing capacity.

Ratings – 6.6V maximum, 0.24A maximum.

HV-aKCEP Series. Platform Scale.

a = Maximum capacity: 15K = 15 kg;
60K = 60 kg;
200K = 220 kg.

HW-aKCEP Series. Platform Scale.

a = Maximum capacity: 10K = 10 kg;
60K = 60 kg;
100K = 100 kg;
200K = 220 kg.

13. Specific Conditions of Use:

1. A portion of the enclosure is non-conducting and, under certain extreme conditions, may generate an ignition-capable level of electrostatic charges. The user shall ensure that the equipment is not installed in a location where it may be subjected to external conditions (such as high-pressure steam) which might cause a build-up of electrostatic charges on non-conducting surfaces. Additionally, cleaning of the equipment should be done only with a damp cloth.
2. The enclosure contains aluminum and is considered a potential risk of ignition by impact or friction. Care must be taken during installation to prevent impact or friction.

14. Test and Assessment Procedure and Conditions:

This Certificate has been issued in accordance with FM Approvals Canadian Certification Scheme.

15. Schedule Drawings

THIS CERTIFICATE MAY ONLY BE REPRODUCED IN ITS ENTIRETY AND WITHOUT CHANGE

FM Approvals LLC. 1151 Boston-Providence Turnpike, Norwood, MA 02062 USA
T: +1 (1) 781 762 4300 F: +1 (1) 781 762 9375 E-mail: information@fmaprovals.com www.fmaprovals.com

SCHEDULE



Canadian Certificate Of Conformity No: FM17CA0089X

A copy of the technical documentation has been kept by FM Approvals.

16. Certificate History

Details of the supplements to this certificate are described below:

Date	Description
12 th January 2018	Original Issue.

THIS CERTIFICATE MAY ONLY BE REPRODUCED IN ITS ENTIRETY AND WITHOUT CHANGE

FM Approvals LLC. 1151 Boston-Providence Turnpike, Norwood, MA 02062 USA
T: +1 (1) 781 762 4300 F: +1 (1) 781 762 9375 E-mail: information@fmaprovals.com www.fmaprovals.com